

ORDER DATE: \_\_\_\_\_

PRESCRIBING PHYSICIAN: \_\_\_\_\_

ACCOUNT CONTACT INFORMATION

Name: \_\_\_\_\_  
Email: \_\_\_\_\_  
Phone: \_\_\_\_\_

SHIPPING PREFERENCE

Ground  
 Next Day  
 2-Day (Standard if no preference is selected)

BILLING INFORMATION

PO #: \_\_\_\_\_ Account #: \_\_\_\_\_  
Address: \_\_\_\_\_  
City: \_\_\_\_\_ State: \_\_\_\_\_ Zip: \_\_\_\_\_

SHIPPING INFORMATION

Company/Account : \_\_\_\_\_  
Address: \_\_\_\_\_  
City: \_\_\_\_\_ State: \_\_\_\_\_ Zip: \_\_\_\_\_

PATIENT INFORMATION

Name: \_\_\_\_\_ Age: \_\_\_\_\_ Gender:  Male  Female  
Email: \_\_\_\_\_ Height: \_\_\_\_\_ (in) Leg:  Left  Right  
Phone: \_\_\_\_\_ Weight: \_\_\_\_\_ (lbs) Patient Diagnosis: \_\_\_\_\_

ORDER INFORMATION

KNEE COMPONENT

Calf Cuff Position:

- Posterior
- Anterior

Hinge Torque:

- Standard
- MAXX, tension lock (+200)

Extension Limits:

- Insert Kit - Adjustability from 5 - 30 degrees at 5 degree increments (+20)
- Built-in Kit - Pre-set to desired angle (+50)

Locking Options:

- Positional Lock for M+L (+50)

Varus Adjustment:

- Velcro Spacer Kit
- Built-in

Valgus Adjustment:

- Velcro Spacer Kit
- Built-in

ANKLE COMPONENT

AFO Type:

- Standard - Thrive, carbon fiber
- Other\* \_\_\_\_\_

\*If Other is selected, you will need to ship your AFO of choice to Icarus for assembly

AFO Size:

- X-Small  Large
- Small  X-Large
- Medium

ADDITIONAL OPTIONS

- Custom Design\* (up to +500)
- Additional Engraving (+50-100)
- Extra Padding (+50)

\*Please include details for any customization in the special instructions box.

ATTACHMENT

- Permanent - Pre-drilled holes with barrel bolts, optionally assembled
- Rapid Detachment - Side lock with snapping caps (+100)

REQUIRED MEASUREMENT


Distance from knee center to bottom of heel/floor \_\_\_\_\_ (in)

SPECIAL INSTRUCTIONS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

## SCANNING INSTRUCTIONS

1. The patient should stand with their leg straight throughout the scan. It is important that they are bearing weight on the leg as well. The scan should start from the inside of the leg with the camera angled so that it is capturing the knee in the center of the screen. Then scan should capture the leg from the foot to 8" above the patella.
2. Press the circular button located at the bottom of the screen and wait until the countdown has finished before moving the phone around the leg from the inside toward the outside in a slow, steady sweeping motion, keeping the phone level as you go.
3. Ideally, you want to capture a full 360 degrees of the leg being scanned; however, 270 degrees will typically provide enough information to accurately depict the curvature of the leg. When you are confident that you have captured at least 270 degrees of their leg, you can end the scan by pressing the circular button at the bottom of the screen.
4. Draw a **“+”** on the center of the knee cap and tap “take picture.” Take the photo straight on and confirm the image once you have a clear photo.
5. Enter the patient’s first name, last name, and email address. Tap “Send Scan” to submit.



Scan to download the  
Icarus Medical App

